



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: BAS717F			
ANILINE HPLC MOBILE PHASE			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON FILE	YES		YES
Sample ID			
Tested 05/31/24			
Tester JRIO			
Recorded 05/31/24			
CALCULATED GPL	1.9823	1.9818 - 1.9828	1.9823
TETRABUTYLAMMONIUM HYDROG	GPL		
Sample ID			
Tested 05/31/24			
Tester JRIO			
Recorded 05/31/24			
CALCULATED V/V% ACETONITRILE	41.0	40.8 - 41.3	41.0
	V/V%		
Sample ID			
Tested 05/31/24			
Tester JRIO			
Recorded 05/31/24			
FILTERED THROUGH 0.2 UM FILTER	YES		YES
	YES		
Sample ID			
Tested 05/31/24			
Tester JRIO			
Recorded 05/31/24			
APPEARANCE AND COLOR	CLEAR LIQUID APPEARANCE		PASS
Sample ID			
Tested 05/31/24			
Tester JRIO			
Recorded 05/31/24			
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		T-222,M115,GC,M117
Sample ID			
Tested 05/31/24			
Tester JRIO			
Recorded 05/31/24			
INITIAL PH OF SOLUTION	6.500	6.480 - 6.520	6.507
	PH		
Sample ID			
Tested 05/31/24			
Tester JRIO			
Recorded 05/31/24			
FINAL PH	6.500	6.300 - 6.900	6.408
	PH		
Sample ID			
Tested 05/31/24			
Tester JRIO			
Recorded 05/31/24			



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

SHELF LIFE OF PRODUCT	6 MONTHS TIME	6 MONTHS
Sample ID		
Tested	05/31/24	
Tester	JRIO	
Recorded	05/31/24	
EXPIRATION DATE	MM/DD/YYYY MM/DD/YYYY	12/30/2024
Sample ID		
Tested	05/31/24	
Tester	JRIO	
Recorded	05/31/24	
Lot #	4053003	
Made	05/31/24	

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager