



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Test	Target/UOM	Range	Result
Packaged Product: DC664F			
	HYDROCHLORIC ACID 4.0 NORMAL		
PREPARED TO FORMULATION ON FILE	YES		YES
Sample ID			
Tested 05/29/24			
Tester ENOR			
Recorded 05/29/24			
SAMPLE COLLECTION & HANDLING PER LAB040	YES		YES
Sample ID			
Tested 05/29/24			
Tester ENOR			
Recorded 05/29/24			
DATE OF SAMPLING FOR TEST PURPOSES	MM/DD/YYYY		05/29/2024
Sample ID			
Tested 05/29/24			
Tester ENOR			
Recorded 05/29/24			
AMBIENT TEMPERATURE DURING SAMPLING DEGC	25.0 DEG C	10.0 - 35.0	20.1
Sample ID			
Tested 05/29/24			
Tester ENOR			
Recorded 05/29/24			
DATE OF TESTING OF SAMPLE	MM/DD/YYYY MM/DD/YYYY		05/29/2024
Sample ID			
Tested 05/29/24			
Tester ENOR			
Recorded 05/29/24			
AMBIENT TEST TEMPERATURE DEGC	20.0 DEG C	15.0 - 30.0	20.0
Sample ID			
Tested 05/29/24			
Tester ENOR			
Recorded 05/29/24			
Actual Normality	4.0000 EQ/L	3.9980 - 4.0020	4.0005
Sample ID			
Tested 05/29/24			
Tester ENOR			
Recorded 05/29/24			
N.I.S.T. Traceable to SRM 723 THAM	YES YES		YES
Sample ID			
Tested 05/29/24			
Tester ENOR			
Recorded 05/29/24			



Aqua Solutions, Inc.
 Ph: 281-479-2569
 Fax: 281-479-2790
 P.O. Box 70
 Deer Park, TX 77536
 Website: www.aquasolutions.org

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569

Certificate of Analysis / QC Results

Customer PO # :

APPEARANCE AND COLOR

CLEAR LIQUID
 APPEARANCE

PASS

Sample ID

Tested 05/29/24

Tester ENOR

Recorded 05/29/24

INSTRUMENTS USED DURING

INSTRUMENT

T-62,GC,T-204,BID-08446,PID-
 M91

PREPARATION

Sample ID

Tested 05/29/24

Tester ENOR

Recorded 05/29/24

Lot # 4052891

Made 05/29/24

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager