



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: BAS088			
REDOX CALIBRATION STD. 200mV +/-10 GLI#25M2A1001-115 (500ML)			
Test	Target/UOM	Range	Result
MILLIVOLT READING	195 TO 205 MV MV		199.0 MV
Sample ID			
Tested	05/31/24		
Tester	TBUR		
Recorded	05/31/24		
EXPIRATION DATE	MM/DD/YY MM/DD/YYYY		04/30/2025
Sample ID			
Tested	05/31/24		
Tester	TBUR		
Recorded	05/31/24		
MANUFACTURE DATE	MM/DD/YY MM/DD/YYYY		05/01/2024
Sample ID			
Tested	05/31/24		
Tester	TBUR		
Recorded	05/31/24		
SHELF LIFE OF PRODUCT	1 YEAR TIME		1 YEAR
Sample ID			
Tested	05/31/24		
Tester	TBUR		
Recorded	05/31/24		
Lot #	415011		
Made	05/29/24		

Richard [Signature]

Quality Manager