

Page: 1

Date: 05/31/24 at 10:39 AM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569

Certificate of Analysis / QC Results

Customer PO #:

Packaged Product:	0373			
		STANDARD 1,000 PPM AS	NH3 (AMMONIUM CHLORIDE)	
Test		Target/UOM	Range	Result
PREPARED TO FORMU	LATION ON	YES	-	YES
FILE		YES		
Sample ID				
Tested	05/31/24			
Tester	AROS			
Recorded	05/31/24			
PPM NH3, CALCULATE	D AMOUNT	1000	960 - 1040	1017
		PPM		
Sample ID				
Tested	05/31/24			
Tester	AROS			
Recorded	05/31/24			
APPEARANCE AND COI		CLEAR LIQUID		PASS
		APPEARANCE		
Sample ID		7 - 2		
Tested	05/31/24			
Tester	AROS			
Recorded	05/31/24			
INSTRUMENTS USED D		INSTRUMENT		M108,GC,VF3-J15,M107
PREPARATION	Uning	INSTRUMENT		W108,GG,V1 3-313,IW107
Sample ID	05/04/04			
Tested	05/31/24			
Tester	AROS			
Recorded	05/31/24	MATERIAL		00/00/04
EXPIRATION DATE		MM/DD/YY		09/30/24
		MM/DD/YYYY		
Sample ID				
Tested	05/31/24			
Tester	AROS			
Recorded	05/31/24			
SHELF LIFE OF PRODU	СТ	SHELF LIFE		3 MONTHS
		TIME		
Sample ID				
Tested	05/31/24			
Tester	AROS			
Recorded	05/31/24			
Lot # 4053154				
Made 05/31				
IVIAGE 50751	,			



Page: 2

Date: 05/31/24 at 10:39 AM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager