

Page: 1

Date: 06/05/24 at 11:28 AM

## Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Packaged Product	: CIT014	ŀF		
	FREE C	CHLORINE STANDARD 175 PP	M (SODIUM HYPOCHLORITE)	
Test		Target/UOM	Range	Result
PREPARED TO FORMU	JLATION ON	YES		YES
FILE		YES		
Sample ID				
Tested	06/05/24			
Tester	JENC			
Recorded	06/05/24			
PPM CHLORINE (CL2)		175.0	170.0 - 180.0	178.0
		PPM		
Sample ID				
Tested	06/05/24			
Tester	JENC			
Recorded	06/05/24			
APPEARANCE AND CO	LOR	CLEAR TO PALE YELLOW		PASS
		LIQUID		
		APPEARANCE		
Sample ID				
Tested	06/05/24			
Tester	JENC			
Recorded	06/05/24			
INSTRUMENTS USED I	DURING	INSTRUMENT		M103,M100,T-161,BID06419,PI
PREPARATION				DP54
Sample ID				
Tested	06/05/24			
Tester	JENC			
Recorded	06/05/24			
Actual Density		0.99966	0.99944 - 0.99988	0.99959
		GM/ML		
Sample ID				
Tested	06/05/24			
Tester	JENC			
Recorded	06/05/24			
EXPIRATION DATE		MM/DD/YYYY		08/10/2024
		MM/DD/YYYY		
Sample ID				
Tested	06/05/24			
Tester	JENC			
Recorded	06/05/24			
Lot # 40604				
Made 06/05/24				
wade 00703	) / 4 t			



Page: 2

Date: 06/05/24 at 11:28 AM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

**Quality Manager**