



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: 8810			
SODIUM THIOSULFATE 0.1 NORMAL			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON FILE	YES		YES
Sample ID	YES		
Tested 06/04/24			
Tester SHAR			
Recorded 06/04/24			
PH OF PRODUCT/SOLUTION	7.000	6.000 - 12.500	10.188
	PH		
Sample ID			
Tested 06/04/24			
Tester SHAR			
Recorded 06/04/24			
Actual Normality	0.1000	0.0995 - 0.1005	0.1005
	EQ/L		
Sample ID			
Tested 06/04/24			
Tester SHAR			
Recorded 06/04/24			
QCSP10:STANDARDIZATION OF SODIUM THIOSUL	YES		YES
	QCSP10		
Sample ID			
Tested 06/04/24			
Tester SHAR			
Recorded 06/04/24			
QCSP35:STANDARDIZATION OF SODIUM THIOSUL	YES		YES
	QCSP35		
Sample ID			
Tested 06/04/24			
Tester SHAR			
Recorded 06/04/24			
APPEARANCE AND COLOR	CLEAR WATER WHITE		PASS
	LIQUID		
	APPEARANCE		
Sample ID			
Tested 06/04/24			
Tester SHAR			
Recorded 06/04/24			
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		M87,M112,T-36,BID-079,PID-M 91
Sample ID			
Tested 06/04/24			
Tester SHAR			
Recorded 06/04/24			
DATE OF PREPARATION	MM/DD/YYYY		06/04/2024
Sample ID			
Tested 06/04/24			
Tester SHAR			
Recorded 06/04/24			



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

EXPIRATION DATE	MM/DD/YY MM/DD/YYYY		12/30/34
Sample ID			
Tested	06/04/24		
Tester	SHAR		
Recorded	06/04/24		
SHELF LIFE	6 MONTHS TIME		6 MONTHS
Sample ID			
Tested	06/04/24		
Tester	SHAR		
Recorded	06/04/24		
TRACEABLE TO NIST SRM 136	YES YES		YES
Sample ID			
Tested	06/04/24		
Tester	SHAR		
Recorded	06/04/24		
Actual Density	1.01127 GM/ML	1.01080 - 1.01173	1.01137
Sample ID			
Tested	06/04/24		
Tester	SHAR		
Recorded	06/04/24		
Lot #	4053051		
Made	06/04/24		

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.


Quality Manager