

## Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

## Customer PO # :

Packaged Produ	ct: 6928			
i uonugou i rouu	•••	CIVITY STANDARD 20.0 uMHC	S	
Test		Target/UOM	Range	Result
PREPARED TO FOR	MULATION ON	YES		YES
FILE		YES		
Sample	ID			
Tested	06/04/24			
Tester	RSHE			
Recorde				
ACTUAL CONDUCTIVITY @25°C		20.0 US/CM	19.9 - 20.1	20.0
Sample	ID			
Tested	06/04/24			
Tester	RSHE			
Recorde				
QCSP30:CONDUCTIVITY		YES		YES
STD&MEASUREMEN		QCSP30		
Sample				
Tested	06/04/24			
Tester	RSHE			
Recorde				N/FO
TRACEABLE TO N.I.S	S.T. SRM 3191	YES		YES
		YES		
Sample				
Tested	06/04/24			
Tester	RSHE			
Recorde		CLEAR WATER WHITE		PASS
APPEARANCE AND	LOLOR			FA35
		APPEARANCE		
Sample	חו	AFFEANANCE		
Tested	06/04/24			
Tester	RSHE			
Recorde				
INSTRUMENTS USE		INSTRUMENT		T-234
PREPARATION				
Sample	D			
Tested	06/04/24			
Tester	RSHE			
Recorde				
SHELF LIFE OF PRO	DUCT	6 MONTHS		6 MONTHS
		TIME		
Sample	ID			
Tested	06/04/24			
Tester	RSHE			
Recorde	d 06/04/24			
EXPIRATION DATE		MM/DD/YYYY MM/DD/YYYY		12/30/2024
Sample	ID			
Tested	06/04/24			
Tester	RSHE			
Recorde	d 06/04/24			



## Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO # :

Lot #	4060404
Made	06/04/24

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

chard St

**Quality Manager**