



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

| Packaged Product: ND497F | | | |
|--|-------------------|-------------------|---------|
| Density Standard Ultra Pure Water | | | |
| Test | Target/UOM | Range | Result |
| PREPARED TO FORMULATION ON FILE | YES | | YES |
| Sample ID | | | |
| Tested 06/05/24 | | | |
| Tester JTRA | | | |
| Recorded 06/05/24 | | | |
| DENSITY @ 10°C | 0.99970 GM/CM3 | 0.99968 - 0.99972 | 0.99970 |
| Sample ID | | | |
| Tested 06/05/24 | | | |
| Tester JTRA | | | |
| Recorded 06/05/24 | | | |
| DENSITY @ 15°C | 0.99910 GM/CM3 | 0.99908 - 0.99912 | 0.99910 |
| Sample ID | | | |
| Tested 06/05/24 | | | |
| Tester JTRA | | | |
| Recorded 06/05/24 | | | |
| DENSITY @ 15.56°C (60°F) | 0.99901 GM/ML | 0.99899 - 0.99903 | 0.99901 |
| Sample ID | | | |
| Tested 06/05/24 | | | |
| Tester JTRA | | | |
| Recorded 06/05/24 | | | |
| DENSITY @ 20°C | 0.99820 GM/CM3 | 0.99818 - 0.99822 | 0.99820 |
| Sample ID | | | |
| Tested 06/05/24 | | | |
| Tester JTRA | | | |
| Recorded 06/05/24 | | | |
| DENSITY @ 25°C | 0.99704 GM/CM3 | 0.99702 - 0.99706 | 0.99704 |
| Sample ID | | | |
| Tested 06/05/24 | | | |
| Tester JTRA | | | |
| Recorded 06/05/24 | | | |
| DENSITY @ 30°C | 0.99565 GM/CM3 | 0.99563 - 0.99567 | 0.99565 |
| Sample ID | | | |
| Tested 06/05/24 | | | |
| Tester JTRA | | | |
| Recorded 06/05/24 | | | |
| DENSITY @ 35°C | 0.99403 GM/CM3 | 0.99401 - 0.99405 | 0.99403 |
| Sample ID | | | |
| Tested 06/05/24 | | | |
| Tester JTRA | | | |
| Recorded 06/05/24 | | | |



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569

Certificate of Analysis / QC Results

Customer PO # :

| | | | |
|-----------------------------|----------------------------|-------------------|----------|
| DENSITY @ 40°C | 0.99221 GM/CM3 | 0.99219 - 0.99223 | 0.99221 |
| Sample ID | | | |
| Tested | 06/05/24 | | |
| Tester | JTRA | | |
| Recorded | 06/05/24 | | |
| DENSITY @ 45°C | 0.99021 GM/CM3 | 0.99019 - 0.99023 | 0.99021 |
| Sample ID | | | |
| Tested | 06/05/24 | | |
| Tester | JTRA | | |
| Recorded | 06/05/24 | | |
| DENSITY @ 50°C | 0.98803 GM/CM3 | 0.98801 - 0.98805 | 0.98803 |
| Sample ID | | | |
| Tested | 06/05/24 | | |
| Tester | JTRA | | |
| Recorded | 06/05/24 | | |
| APPEARANCE AND COLOR | CLEAR LIQUID APPEARANCE | | PASS |
| Sample ID | | | |
| Tested | 06/05/24 | | |
| Tester | JTRA | | |
| Recorded | 06/05/24 | | |
| EXPIRATION DATE | MM/DD/YY MM/DD/YYYY | | 12/30/25 |
| Sample ID | | | |
| Tested | 06/05/24 | | |
| Tester | JTRA | | |
| Recorded | 06/05/24 | | |
| Lot # | 4060482 | | |
| Made | 06/05/24 | | |

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager