



**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

<b>Packaged Product:</b> 9026			
SULFURIC ACID 0.03 NORMAL			
Test	Target/UOM	Range	Result
Prepared to formulation on File	Yes		YES
	YES		
Sample ID			
Tested	06/04/24		
Tester	AFRE		
Recorded	06/04/24		
<b>SAMPLE COLLECTION &amp; HANDLING</b>	YES		YES
<b>PER LAB040</b>	YES		
Sample ID			
Tested	06/04/24		
Tester	AFRE		
Recorded	06/04/24		
<b>DATE OF SAMPLING FOR TEST</b>	MM/DD/YYYY		06/03/2024
<b>PURPOSES</b>	MM/DD/YYYY		
Sample ID			
Tested	06/04/24		
Tester	AFRE		
Recorded	06/04/24		
<b>AMBIENT TEMPERATURE DURING</b>	25.0	10.0 - 35.0	20.0
<b>SAMPLING DEGC</b>	DEG C		
Sample ID			
Tested	06/04/24		
Tester	AFRE		
Recorded	06/04/24		
<b>DATE OF TESTING OF SAMPLE</b>	MM/DD/YYYY		06/03/2024
	MM/DD/YYYY		
Sample ID			
Tested	06/04/24		
Tester	AFRE		
Recorded	06/04/24		
<b>AMBIENT TEST TEMPERATURE</b>	20.0	15.0 - 30.0	20.0
<b>DEG C</b>	DEG C		
Sample ID			
Tested	06/04/24		
Tester	AFRE		
Recorded	06/04/24		
<b>Actual Normality</b>	0.0300	0.0295 - 0.0305	0.0305
	EQ/L		
Sample ID			
Tested	06/04/24		
Tester	AFRE		
Recorded	06/04/24		
<b>N.I.S.T. Traceable to SRM 723 THAM</b>	YES		YES
	YES		
Sample ID			
Tested	06/04/24		
Tester	AFRE		
Recorded	06/04/24		



**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

<b>QCSP5: STANDARDIZATION OF AQUEOUS ACIDS</b>	YES METHOD	YES
Sample ID		
Tested 06/04/24		
Tester AFRE		
Recorded 06/04/24		
<b>APPEARANCE AND COLOR</b>	CLEAR WATER WHITE LIQUID APPEARANCE	PASS
Sample ID		
Tested 06/04/24		
Tester AFRE		
Recorded 06/04/24		
<b>INSTRUMENTS USED DURING PREPARATION</b>	INSTRUMENT	GC,VF4-I6
Sample ID		
Tested 06/04/24		
Tester AFRE		
Recorded 06/04/24		
<b>EXPIRATION DATE</b>	MM/DD/YYYY MM/DD/YYYY	06/30/2025
Sample ID		
Tested 06/04/24		
Tester AFRE		
Recorded 06/04/24		
<b>Lot #</b> 4053142		
<b>Made</b> 06/04/24		

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.



**Quality Manager**