



**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

Test	Target/UOM	Range	Result
<b>Packaged Product:</b> WES035F			
	Water Standard (8 Component Solution in 1% HNO3)		
<b>PREPARED TO FORMULATION ON FILE</b>	YES		YES
Sample ID			
Tested 06/04/24			
Tester AROS			
Recorded 06/04/24			
<b>CALCULATED PPM CALCIUM</b>	15.00	14.90 - 15.10	15.00
	PPM		
Sample ID			
Tested 06/04/24			
Tester AROS			
Recorded 06/04/24			
<b>CALCULATED PPM MAGNESIUM (MG)</b>	15.00	14.90 - 15.10	15.00
	PPM		
Sample ID			
Tested 06/04/24			
Tester AROS			
Recorded 06/04/24			
<b>CALCULATED PPM IRON (FE)</b>	1.00	0.99 - 1.01	1.00
	PPM		
Sample ID			
Tested 06/04/24			
Tester AROS			
Recorded 06/04/24			
<b>CALCULATED MG/L SILICON (SI)</b>	1.00	0.99 - 1.01	1.00
	MG/L		
Sample ID			
Tested 06/05/24			
Tester AROS			
Recorded 06/05/24			
<b>CALCULATED PPM PHOSPHATE (PO4)</b>	15.00	14.90 - 15.10	15.00
	PPM (MG/L)		
Sample ID			
Tested 06/05/24			
Tester AROS			
Recorded 06/05/24			
<b>CALCULATED PPM MOLYBDENUM (MO)</b>	15.00	14.90 - 15.10	15.00
	PPM		
Sample ID			
Tested 06/05/24			
Tester AROS			
Recorded 06/05/24			
<b>CALCULATED PPM CADMIUM</b>	1.00	0.99 - 1.01	1.00
	MG/L		
Sample ID			
Tested 06/05/24			
Tester AROS			
Recorded 06/05/24			



Aqua Solutions, Inc.  
 Ph: 281-479-2569  
 Fax: 281-479-2790  
 P.O. Box 70  
 Deer Park, TX 77536  
 Website: [www.aquasolutions.org](http://www.aquasolutions.org)

**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

<b>CALCULATED PPM CHROMIUM</b>	1.00	0.99 - 1.01	1.00
	PPM		
Sample ID			
Tested	06/05/24		
Tester	AROS		
Recorded	06/05/24		
<b>APPEARANCE AND COLOR</b>	PASS		PASS
	APPEARANCE		
Sample ID			
Tested	06/05/24		
Tester	AROS		
Recorded	06/05/24		
<b>INSTRUMENTS USED DURING PREPARATION</b>	INSTRUMENT		P17,P184,BOTTLE,GP,VF1-Z3 5
Sample ID			
Tested	06/05/24		
Tester	AROS		
Recorded	06/05/24		
<b>EXPIRATION DATE</b>	MM/DD/YYYY		06/30/2025
	MM/DD/YYYY		
Sample ID			
Tested	06/05/24		
Tester	AROS		
Recorded	06/05/24		
<b>Lot #</b>	4060467		
<b>Made</b>	06/05/24		

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

  
**Quality Manager**