



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: BAS732F			
SODIUM HYDROXIDE 1.0 NORMAL, LOW CARBONATE			
Test	Target/UOM	Range	Result
EXPIRATION DATE	MM/DD/YYYY		06/30/2026
	MM/DD/YYYY		
Sample ID			
Tested 06/06/24			
Tester JENC			
Recorded 06/06/24			
PREPARED TO FORMULATION ON FILE	YES		YES
	YES		
Sample ID			
Tested 06/03/24			
Tester JENC			
Recorded 06/03/24			
SAMPLE COLLECTION & HANDLING PER LAB040	YES		YES
	YES		
Sample ID			
Tested 06/03/24			
Tester JENC			
Recorded 06/03/24			
DATE OF SAMPLING FOR TEST PURPOSES	MM/DD/YYYY		06/03/2024
	MM/DD/YYYY		
Sample ID			
Tested 06/03/24			
Tester JENC			
Recorded 06/03/24			
AMBIENT TEMPERATURE DURING SAMPLING DEGC	25.0	10.0 - 35.0	19.7
	DEG C		
Sample ID			
Tested 06/03/24			
Tester JENC			
Recorded 06/03/24			
DATE OF TESTING OF SAMPLE	MM/DD/YYYY		06/03/2024
	MM/DD/YYYY		
Sample ID			
Tested 06/03/24			
Tester JENC			
Recorded 06/03/24			
AMBIENT TEST TEMPERATURE DEGC	20.0	15.0 - 30.0	19.7
	DEG C		
Sample ID			
Tested 06/03/24			
Tester JENC			
Recorded 06/03/24			
NORMALITY TO 4 DEC PLACES	1.0000	0.9995 - 1.0005	1.0003
	EQUIV/LITER		
Sample ID			
Tested 06/03/24			
Tester JENC			
Recorded 06/03/24			



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

PPM INORGANIC CARBON	10.00 PPM	0.00 - 22.00	18.00
Sample ID			
Tested	06/03/24		
Tester	JENC		
Recorded	06/03/24		
Actual Density	1.04101 GM/ML	1.04085 - 1.04117	1.04103
Sample ID			
Tested	06/03/24		
Tester	JENC		
Recorded	06/03/24		
N.I.S.T. Traceable to SRM 84	YES YES		YES
Sample ID			
Tested	06/03/24		
Tester	JENC		
Recorded	06/03/24		
APPEARANCE AND COLOR	CLEAR LIQUID APPEARANCE		PASS
Sample ID			
Tested	06/03/24		
Tester	JENC		
Recorded	06/03/24		
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		M87,T-45,BID-02259,PID-M91
Sample ID			
Tested	06/03/24		
Tester	JENC		
Recorded	06/03/24		
SHELF LIFE OF PRODUCT	2 YEARS TIME		2 YEARS
Sample ID			
Tested	06/03/24		
Tester	JENC		
Recorded	06/03/24		
Lot #	4053160		
Made	06/03/24		

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.



Aqua Solutions, Inc.
Ph: 281-479-2569
Fax: 281-479-2790
P.O. Box 70
Deer Park, TX 77536
Website: www.aquasolutions.org

Page: 3

Date: 06/06/24 at 8:19 AM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager