

## Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

## Customer PO # :

Packaged Product: 4360			
i uokugeu i iouuoti	HLORIC ACID 5.0 NORMAL		
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON	YES	nango	YES
FILE	YES		
Sample ID			
Tested 06/05/24			
Tester AFRE			
Recorded 06/05/24			
SAMPLE COLLECTION & HANDLING	YES		YES
PER LAB040	YES		
Sample ID			
Tested 06/05/24			
Tester AFRE			
Recorded 06/05/24			
DATE OF SAMPLING FOR TEST	MM/DD/YYYY		06/05/2024
PURPOSES	MM/DD/YYYY		
Sample ID			
Tested 06/05/24			
Tester AFRE			
Recorded 06/05/24			
AMBIENT TEMPERATURE DURING	25.0	10.0 - 35.0	20.0
SAMPLING DEGC	DEG C		
Sample ID			
Tested 06/05/24			
Tester AFRE			
Recorded 06/05/24			
DATE OF TESTING OF SAMPLE	MM/DD/YYYY MM/DD/YYYY		06/05/2024
Sample ID			
Tested 06/05/24			
Tester AFRE			
Recorded 06/05/24			
AMBIENT TEST TEMPERATURE	20.0	15.0 - 30.0	20.0
DEG C	DEG C		
Sample ID			
Tested 06/05/24			
Tester AFRE			
Recorded 06/05/24			
APPEARANCE AND COLOR	CLEAR LIQUID APPEARANCE		PASS
Sample ID			
Tested 06/05/24			
Tester AFRE			
Recorded 06/05/24			
QCSP5: STANDARDIZATION OF	YES		YES
AQUEOUS ACIDS	METHOD		
Sample ID			
Tested 06/05/24			
Tester AFRE			
Recorded 06/05/24			



Page: 2 Date: 06/05/24 at 11:30 AM

## Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO # :

INSTRUMENTS USED D PREPARATION	URING	INSTRUMENT		M87,T-230,M55,BID00488,PIE M91
Sample ID Tested Tester Recorded	06/05/24 AFRE 06/05/24			
Actual Density		1.08137	1.08053 - 1.08221	1.08055
		GM/ML		
Sample ID				
Tested	06/05/24			
Tester	AFRE			
Recorded NORMALITY TO 4 DEC	06/05/24	5.0000	4.0800 5.0000	4.9870
NORMALITY TO 4 DEC	PLACES		4.9800 - 5.0200	4.9870
Comple ID		EQUIV/LITER		
Sample ID Tested	00/05/04			
	06/05/24			
Tester	AFRE			
Recorded EXPIRATION DATE	06/05/24	MM/DD/YY		06/30/26
EXPIRATION DATE		MM/DD/YYYY		00/30/20
Sample ID				
Tested	06/05/04			
Tester	06/05/24 AFRE			
Recorded	06/05/24			
COLOR	00/03/24	< 10		5
COLON		PT-CO		3
Sample ID		11-66		
Tested	06/05/24			
Tester	AFRE			
Recorded	06/05/24			
N.I.S.T. Traceable to SR		YES		YES
		YES		120
Sample ID		0		
Tested	06/05/24			
Tester	AFRE			
Recorded	06/05/24			
SHELF LIFE OF PRODU		2 YEARS		2 YEARS
	-	TIME		
Sample ID				
Tested	06/05/24			
Tester	AFRE			
Recorded	06/05/24			
Lot # 40603				
Made 06/05	/ 24			



## Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results Customer PO # :

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

1 chard Stop

**Quality Manager**