



**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

<b>Packaged Product:</b> ERL114AF			
POTASSIUM HYDROXIDE 0.1 NORMAL IN IPA			
Test	Target/UOM	Range	Result
<b>PREPARED TO FORMULATION ON FILE</b>	YES		YES
Sample ID			
Tested 06/06/24			
Tester SHAR			
Recorded 06/06/24			
<b>SAMPLE COLLECTION &amp; HANDLING PER LAB040</b>	YES		YES
Sample ID			
Tested 06/06/24			
Tester SHAR			
Recorded 06/06/24			
<b>DATE OF SAMPLING FOR TEST PURPOSES</b>	MM/DD/YYYY		06/06/2024
Sample ID			
Tested 06/06/24			
Tester SHAR			
Recorded 06/06/24			
<b>AMBIENT TEMPERATURE DURING SAMPLING DEGC</b>	25.0	10.0 - 35.0	21.0
Sample ID			
Tested 06/06/24			
Tester SHAR			
Recorded 06/06/24			
<b>DATE OF TESTING OF SAMPLE</b>	MM/DD/YYYY		06/06/2024
Sample ID			
Tested 06/06/24			
Tester SHAR			
Recorded 06/06/24			
<b>AMBIENT TEST TEMPERATURE DEGC</b>	20.0	15.0 - 30.0	21.0
Sample ID			
Tested 06/06/24			
Tester SHAR			
Recorded 06/06/24			
<b>LOT NUMBER OF ORIGINAL MATERIAL</b>	BULK LOT NUMBER		4031808
Sample ID	LOT #		
Tested 06/06/24			
Tester SHAR			
Recorded 06/06/24			
<b>Actual Normality</b>	0.1000	0.0995 - 0.1005	0.0999
Sample ID	EQ/L		
Tested 06/06/24			
Tester SHAR			
Recorded 06/06/24			



**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**

**Certificate of Analysis / QC Results**

Customer PO # :

<b>N.I.S.T. Traceable to SRM 84 (KHP)</b>	YES YES	YES
Sample ID		
Tested	06/06/24	
Tester	SHAR	
Recorded	06/06/24	
<b>APPEARANCE AND COLOR</b>	CLEAR-SLIGHTLY HAZY LIQUID APPEARANCE	PASS
Sample ID		
Tested	06/06/24	
Tester	SHAR	
Recorded	06/06/24	
<b>INSTRUMENTS USED DURING PREPARATION</b>	LAB EQUIPMENT INSTRUMENT	CUBE,M55,BURET:079,M91
Sample ID		
Tested	06/06/24	
Tester	SHAR	
Recorded	06/06/24	
<b>EXPIRATION DATE</b>	MM/DD/YYYY MM/DD/YYYY	06/18/2024
Sample ID		
Tested	06/06/24	
Tester	SHAR	
Recorded	06/06/24	
<b>SHELF LIFE OF PRODUCT</b>	1 WEEK TIME	1 WEEK
Sample ID		
Tested	06/06/24	
Tester	SHAR	
Recorded	06/06/24	
<b>Lot #</b>	4060307	
<b>Made</b>	06/06/24	

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

**Quality Manager**