

Page: 1

Date: 06/13/24 at 4:39 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569

Certificate of Analysis / QC Results Customer PO #:

			045.0		
Packag	ged Product:	2650			
•	•		STANDARD 1000 PPM 1ml	L=1mg CN (MAKE FRESH)	
Test			Target/UOM	Range	Result
PREPAR	ED TO FORMU	LATION ON	YES	-	YES
FILE			YES		
	Sample ID				
	Tested	06/13/24			
	Tester	JENC			
	Recorded	06/13/24			
APPEARANCE AND COLOR		LOR	CLEAR LIQUID		PASS
			APPEARANCE		
	Sample ID				
	Tested	06/13/24			
	Tester	JENC			
	Recorded	06/13/24			
INSTRUMENTS USED DURING			INSTRUMENT		BID02259,PIDP-177,TID41100
PREPAR	RATION				
	Sample ID				
	Tested	06/13/24			
	Tester	JENC			
	Recorded	06/13/24			
ACTUAL PPM			1000.00	995.00 - 1005.00	998.85
			PPM		
	Sample ID				
	Tested	06/13/24			
	Tester	JENC			
	Recorded	06/13/24			
QCSP47:STANDARDIZATION		ATION	YES		YES
VARIOUS CYANIDE			QCSP47		
	Sample ID				
	Tested	06/13/24			
	Tester	JENC			
	Recorded	06/13/24			
EXPIRATION DATE			MM/DD/YY		07/18/24
			MM/DD/YYYY		
	Sample ID				
	Tested	06/13/24			
	Tester	JENC			
	Recorded	06/13/24			
ı	ot # 40611				
	lade 06/13				
	1440 7 - 2 °	•			



Page: 2

Date: 06/13/24 at 4:39 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager