



**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

<b>Packaged Product:</b> 7260			
POTASSIUM HYDROXIDE 0.5 NORMAL IN METHANOL			
Test	Target/UOM	Range	Result
Actual Normality	0.5000 EQ/L	0.4995 - 0.5005	0.5002
Sample ID			
Tested	06/12/24		
Tester	RIRO		
Recorded	06/12/24		
<b>QCSP7: STANDARDIZATION OF AQUEOUS BASES</b>	YES METHOD		YES
Sample ID			
Tested	06/12/24		
Tester	RIRO		
Recorded	06/12/24		
<b>N.I.S.T. Traceable to SRM 919</b>	Yes YES		YES
Sample ID			
Tested	06/12/24		
Tester	RIRO		
Recorded	06/12/24		
<b>PREPARED TO FORMULATION ON FILE</b>	YES YES		YES
Sample ID			
Tested	06/12/24		
Tester	RIRO		
Recorded	06/12/24		
<b>SAMPLE COLLECTION &amp; HANDLING PER LAB040</b>	YES YES		YES
Sample ID			
Tested	06/12/24		
Tester	RIRO		
Recorded	06/12/24		
<b>DATE OF SAMPLING FOR TEST PURPOSES</b>	MM/DD/YYYY MM/DD/YYYY		06/10/2024
Sample ID			
Tested	06/12/24		
Tester	RIRO		
Recorded	06/12/24		
<b>AMBIENT TEMPERATURE DURING SAMPLING DEGC</b>	25.0 DEG C	10.0 - 35.0	21.0
Sample ID			
Tested	06/12/24		
Tester	RIRO		
Recorded	06/12/24		
<b>DATE OF TESTING OF SAMPLE</b>	MM/DD/YYYY MM/DD/YYYY		06/12/2024
Sample ID			
Tested	06/12/24		
Tester	RIRO		
Recorded	06/12/24		



**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**

**Certificate of Analysis / QC Results**

Customer PO # :

<b>AMBIENT TEST TEMPERATURE</b>	20.0	15.0 - 30.0	21.0
<b>DEG C</b>	DEG C		
Sample ID			
Tested	06/12/24		
Tester	RIRO		
Recorded	06/12/24		
<b>APPEARANCE AND COLOR</b>	CLEAR WATER WHITE		PASS
	LIQUID		
	APPEARANCE		
Sample ID			
Tested	06/12/24		
Tester	RIRO		
Recorded	06/12/24		
<b>INSTRUMENTS USED DURING PREPARATION</b>	INSTRUMENT		M103,T-230,M55,BURET:079, M91
Sample ID			
Tested	06/12/24		
Tester	RIRO		
Recorded	06/12/24		
<b>EXPIRATION DATE</b>	MM/DD/YY		06/30/25
	MM/DD/YYYY		
Sample ID			
Tested	06/12/24		
Tester	RIRO		
Recorded	06/12/24		
<b>Actual Density</b>	0.83392	0.83277 - 0.83506	0.83444
	GM/ML		
Sample ID			
Tested	06/12/24		
Tester	RIRO		
Recorded	06/12/24		
<b>Lot #</b>	4060654		
<b>Made</b>	06/12/24		

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

**Quality Manager**