

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO # :

Packaged Product: 6921			
	UCTIVITY STANDARD 10.0 uMHC	S	
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON			YES
FILE	YES		
Sample ID			
Tested 06/14/24			
Tester BPHI			
Recorded 06/14/24			
CONDUCTIVITY @ 25°C 1 DEC	10.0	9.9 - 10.1	10.0
PLACE	US/CM		
Sample ID			
Tested 06/14/24			
Tester BPHI			
Recorded 06/14/24			
QCSP30:CONDUCTIVITY	YES		YES
STD&MEASUREMENT	QCSP30		
Sample ID			
Tested 06/14/24			
Tester BPHI			
Recorded 06/14/24	NEO.		NEO.
TRACEABLE TO N.I.S.T. SRM 3191			YES
	YES		
Sample ID			
Tested 06/14/24			
Tester BPHI			
Recorded 06/14/24	CLEAR WATER WHITE		PASS
APPEARANCE AND COLOR			PASS
	APPEARANCE		
Sample ID			
Tested 06/14/24			
Tester BPHI			
Recorded 06/14/24 INSTRUMENTS USED DURING	INSTRUMENT		T-229,M92
PREPARATION	INSTRUMENT		1-229,10192
Sample ID Tested 06/14/24			
Tester BPHI Recorded 06/14/24			
EXPIRATION DATE	MM/DD/YY		09/30/24
	MM/DD/YYYY		00/00/24
Sample ID			
Tested 06/14/24			
Tester BPHI			
Recorded 06/14/24			
SHELF LIFE OF PRODUCT	SHELF LIFE		3 MONTHS
	TIME		
Sample ID			
Tested 06/14/24			
Tester BPHI			
Recorded 06/14/24			



Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO # :

Lot #	4061433
Made	06/14/24

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

chard Sto

Quality Manager