

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO # :

Packaged Product: 693	35		
	IDUCTIVITY STANDARD 5,000 ul	MHOS	
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION			YES
FILE	YES		
Sample ID			
Tested 06/12/2	4		
Tester RSHE			
Recorded 06/12/2			
ACTUAL CONDUCTIVITY @25°C	5000.0 US/CM	4950.0 - 5050.0	5000.0
Sample ID			
Tested 06/12/2	4		
Tester RSHE			
Recorded 06/12/2			
QCSP30:CONDUCTIVITY	YES		YES
STD&MEASUREMENT	QCSP30		
Sample ID			
Tested 06/12/2	4		
Tester RSHE			
Recorded 06/12/2			
TRACEABLE TO N.I.S.T. SRM 31			YES
	YES		
Sample ID			
Tested 06/12/2	4		
Tester RSHE			
Recorded 06/12/2			
APPEARANCE AND COLOR	CLEAR WATER WHITE		PASS
	LIQUID		
	APPEARANCE		
Sample ID			
Tested 06/12/2	4		
Tester RSHE			
Recorded 06/12/2			T 66.4
INSTRUMENTS USED DURING	INSTRUMENT		T-234
PREPARATION			
Sample ID			
Tested 06/12/2	4		
Tester RSHE			
Recorded 06/12/2			00/00/05
EXPIRATION DATE			06/30/25
	MM/DD/YYYY		
Sample ID			
Tested 06/12/2	4		
Tester RSHE	4		
Recorded 06/12/2 SHELF LIFE OF PRODUCT	4 SHELF LIFE		1 YEAR
	TIME		I TEAR
Sample ID			
Tested 06/12/2	4		
Tester RSHE			
Recorded 06/12/2	4		



Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO # :

Lot #	4061261
Made	06/12/24

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

chard St

Quality Manager