06/10/24

Made

Page: 1

Date: 06/10/24 at 11:39 AM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Packaged Pro	oduct: M-593F			
Sodium 0.1 ppm ICP Standard with 5% Nitric Acid				
Test		Target/UOM	Range	Result
PREPARED TO F	ORMULATION ON	YES	-	YES
FILE		YES		
Sam	ple ID			
Test	ed 06/10/24			
Test	er MMAT			
Rec	orded 06/10/24			
APPEARANCE AND COLOR		CLEAR WATER WHITE		PASS
		APPEARANCE		
Sam	ple ID			
Test	ed 06/10/24			
Test	er MMAT			
Rec	orded 06/10/24			
INSTRUMENTS U	JSED DURING	INSTRUMENT		P83,GC,VF500-X41
PREPARATION				
Sam	ple ID			
Test	ed 06/10/24			
Test	er MMAT			
Rec	orded 06/10/24			
EXPIRATION DATE		MM/DD/YY		03/30/25
		MM/DD/YYYY		
Sam	ple ID			
Test	•			
Test	er MMAT			
Rec	orded 06/10/24			
PPM SODIUM TO	3 DECIMAL	0.100	0.095 - 0.105	0.100
PLACES		PPM		
Sam	ple ID			
Test	•			
Test	er MMAT			
Rec	orded 06/10/24			
l ot # 4	1060674			
	06/10/04			

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

bulant I