



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

| Packaged Product: ERL125F | | | |
|--|------------|-----------------|------------|
| SILVER NITRATE 0.1 NORMAL IN ISOPROPANOL | | | |
| Test | Target/UOM | Range | Result |
| PREPARED TO FORMULATION ON FILE | YES | | YES |
| Sample ID | YES | | |
| Tested 06/18/24 | | | |
| Tester SHAR | | | |
| Recorded 06/18/24 | | | |
| SAMPLE COLLECTION & HANDLING PER LAB040 | YES | | YES |
| Sample ID | YES | | |
| Tested 06/18/24 | | | |
| Tester SHAR | | | |
| Recorded 06/18/24 | | | |
| DATE OF SAMPLING FOR TEST PURPOSES | MM/DD/YYYY | | 06/18/2024 |
| Sample ID | MM/DD/YYYY | | |
| Tested 06/18/24 | | | |
| Tester SHAR | | | |
| Recorded 06/18/24 | | | |
| AMBIENT TEMPERATURE DURING SAMPLING DEGC | 25.0 | 10.0 - 35.0 | 21.0 |
| Sample ID | DEG C | | |
| Tested 06/18/24 | | | |
| Tester SHAR | | | |
| Recorded 06/18/24 | | | |
| DATE OF TESTING OF SAMPLE | MM/DD/YYYY | | 06/18/2024 |
| Sample ID | MM/DD/YYYY | | |
| Tested 06/18/24 | | | |
| Tester SHAR | | | |
| Recorded 06/18/24 | | | |
| AMBIENT TEST TEMPERATURE DEGC | 20.0 | 15.0 - 30.0 | 21.0 |
| Sample ID | DEG C | | |
| Tested 06/18/24 | | | |
| Tester SHAR | | | |
| Recorded 06/18/24 | | | |
| Actual Normality | 0.1000 | 0.0995 - 0.1005 | 0.1016 |
| Sample ID | EQ/L | | |
| Tested 06/18/24 | | | |
| Tester SHAR | | | |
| Recorded 06/18/24 | | | |
| QCSP53: SILVER NITRATE STD ON AUTO-TITRA | YES | | YES |
| Sample ID | METHOD | | |
| Tested 06/18/24 | | | |
| Tester SHAR | | | |
| Recorded 06/18/24 | | | |



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569

Certificate of Analysis / QC Results

Customer PO # :

| | | | |
|--|--------------|-------------------|----------------------------|
| N.I.S.T. Traceable to SRM 919 NaCl | YES | | YES |
| | YES | | |
| Sample ID | | | |
| Tested | 06/18/24 | | |
| Tester | SHAR | | |
| Recorded | 06/18/24 | | |
| PREPARED TO ASTM D3227 | YES | | YES |
| | YES | | |
| Sample ID | | | |
| Tested | 06/18/24 | | |
| Tester | SHAR | | |
| Recorded | 06/18/24 | | |
| Actual Density | 0.82561 | 0.82176 - 0.82947 | 0.82657 |
| | GM/ML | | |
| Sample ID | | | |
| Tested | 06/18/24 | | |
| Tester | SHAR | | |
| Recorded | 06/18/24 | | |
| APPEARANCE AND COLOR | CLEAR LIQUID | | PASS |
| | APPEARANCE | | |
| Sample ID | | | |
| Tested | 06/18/24 | | |
| Tester | SHAR | | |
| Recorded | 06/18/24 | | |
| INSTRUMENTS USED DURING PREPARATION | INSTRUMENT | | M55,BIDAT2,PIDP18,TID41124 |
| Sample ID | | | |
| Tested | 06/18/24 | | |
| Tester | SHAR | | |
| Recorded | 06/18/24 | | |
| LOT NUMBER OF ORIGINAL MATERIAL | LOT # | | 4040108 |
| Sample ID | | | |
| Tested | 06/18/24 | | |
| Tester | SHAR | | |
| Recorded | 06/18/24 | | |
| EXPIRATION DATE | MM/DD/YY | | 07/02/24 |
| | MM/DD/YYYY | | |
| Sample ID | | | |
| Tested | 06/18/24 | | |
| Tester | SHAR | | |
| Recorded | 06/18/24 | | |
| SHELF LIFE OF PRODUCT | 1 WEEK | | 1 WEEK |
| | TIME | | |
| Sample ID | | | |
| Tested | 06/18/24 | | |
| Tester | SHAR | | |
| Recorded | 06/18/24 | | |

Lot # 4061709
Made 06/18/24



Aqua Solutions, Inc.
Ph: 281-479-2569
Fax: 281-479-2790
P.O. Box 70
Deer Park, TX 77536
Website: www.aquasolutions.org

Page: 3

Date: 06/18/24 at 9:48 AM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager