

Page: 1

Date: 06/17/24 at 11:48 AM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569

Certificate of Analysis / QC Results

Customer PO #:

Customer PO # :					
Package	ed Product:	I7600			
			PYL ALCOHOL REAGENT ACS		
Test			Target/UOM	Range	Result
ACETONE	%		<=0.002	- Harige	<0.002%
	- ,-		%		
	Sample ID				
	Tested	06/17/24			
	Tester	TBUR			
	Recorded	06/17/24			
ASSAY OF MATERIAL			>= 99.5%		99.96%
			%		
	Sample ID				
	Tested	06/17/24			
	Tester	TBUR			
	Recorded	06/17/24			
COLOR			MAX 10 APHA		<10 APHA
			PT-CO		
	Sample ID				
	Tested	06/17/24			
	Tester	TBUR			
	Recorded	06/17/24			
SOLUBILI		00,, = .	PASSES TEST		PASSES TEST
			PASS		
	Sample ID				
	Tested	06/17/24			
	Tester	TBUR			
	Recorded	06/17/24			
WATER CONTENT		00/1//21	MAX 0.2%		0.021%
			%		
	Sample ID				
	Tested	06/17/24			
	Tester	TBUR			
	Recorded	06/17/24			
RESIDUE	AFTER EVAP		<= 0.001%		<0.0001%
	Sample ID				
	Tested	06/17/24			
	Tester	TBUR			
	Recorded	06/17/24			
CARBONY		-0	<= 0.002		<0.002%
-	NALDEHYDE9	6	WT%		10100=70
	Sample ID	-	**		
	Tested	06/17/24			
	Tester	TBUR			
	Recorded	06/17/24			
TITRABLE		30/11/LT	MAX 0.0001 MEQ/G		0.00005 MEQ/G
			T ACID		5.50000 WEG/ G
	Sample ID		1 AOID		
	Tested	06/17/24			
	Tester	TBUR			
	Recorded	06/17/24			

Page: 2

Date: 06/17/24 at 11:48 AM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

TITRABLE BASE MAX 0.0001 MEQ/G <0.00001 MEQ/G

T BASE

Sample ID

Tested 06/17/24
Tester TBUR
Recorded 06/17/24

MANUFACTURE DATE MM/DD/YY 06/03/2024

MM/DD/YYYY

Sample ID

Tested 06/17/24 Tester TBUR Recorded 06/17/24

Lot # 416605 Made 06/14/24

Quality Manager