

Page: 1

Date: 06/18/24 at 2:14 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Packaged Product				
	CHLORID	E STANDARD 100.0 PPM w/	/w (1ml=0.1mg Cl-)	
Test		Target/UOM	Range	Result
PREPARED TO FORM	ULATION ON	YES		YES
FILE		YES		
Sample ID)			
Tested	06/18/24			
Tester	MMAT			
Recorded	06/18/24			
CHLORIDE PPM		100.0	95.0 - 105.0	95.8
		PPM		
Sample ID	1			
Tested	06/18/24			
Tester	MMAT			
Recorded	06/18/24			
APPEARANCE AND COLOR		CLEAR LIQUID		PASS
	- 	APPEARANCE		. , .55
Sample ID	1			
Tested	06/18/24			
Tester	MMAT			
Recorded	06/18/24			
		YES		YES
QCSP74:MEASURING CHLORIDE IN WATER		QCSP74		123
		QC3F74		
Sample ID				
Tested	06/18/24			
Tester	MMAT			
Recorded	06/18/24			
NSTRUMENTS USED	DURING	INSTRUMENT		M108,VF4-N16
PREPARATION				
Sample ID)			
Tested	06/18/24			
Tester	MMAT			
Recorded	06/18/24			
EXPIRATION DATE		MM/DD/YY		12/30/24
		MM/DD/YYYY		
Sample ID	1			
Tested	06/18/24			
Tester	MMAT			
Recorded	06/18/24			
SHELF LIFE OF PROD		6 MONTHS		6 MONTHS
		TIME		
Sample ID)			
Tested	06/18/24			
Tester	MMAT			
Recorded	06/18/24			
Lot # 4061				
Made 06/1	8/24			



Page: 2

Date: 06/18/24 at 2:14 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager