Page: 1

Date: 06/18/24 at 2:10 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Ammonia as Nitrogen Standard 2.0 ppm Aqueous				Solution	
Test		Target/UOM	Range	Result	
PREPARE	D TO FORMULATION C			YES	
FILE		YES			
	Sample ID				
	Tested 06/18/24				
	Tester MMAT				
	Recorded 06/18/24				
PPM NITROGEN 2 DECIMAL		2.00	1.80 - 2.20	1.86	
PLACES		PPM			
	Sample ID				
	Tested 06/18/24				
	Tester MMAT				
	Recorded 06/18/24				
APPEARANCE AND COLOR		CLEAR WATER WHITE		PASS	
		APPEARANCE			
	Sample ID				
	Tested 06/18/24				
	Tester MMAT				
	Recorded 06/18/24				
NSTRUM	ENTS USED DURING	INSTRUMENT		P166,VF2-A13	
PREPARA	ATION				
	Sample ID				
	Tested 06/18/24				
	Tester MMAT				
	Recorded 06/18/24				
EXPIRATION DATE		MM/DD/YYYY		09/30/2024	
		MM/DD/YYYY			
	Sample ID				
	Tested 06/18/24				
	Tester MMAT				
	Recorded 06/18/24				

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

huland I