

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO # :

Packaged Prod	luct: 6490			
-	NITROG	EN STANDARD 1000 PPM (1M)	L=1MG N), CERTIFIED	
Test		Target/UOM	Range	Result
PREPARED TO FORMULATION ON		YES	V	YES
FILE		YES		
Sampl	le ID			
Tested	d 06/20/24			
Tester	· MMAT			
Record	ded 06/20/24			
PPM NITROGEN NO DECIAML		1000	980 - 1020	982
PLACES		PPM		
Sampl	le ID			
Tested				
Tester				
Record	ded 06/20/24			
APPEARANCE AND COLOR		CLEAR WATER WHITE		PASS
		LIQUID		
		APPEARANCE		
Sampl	le ID			
Tested				
Tester				
Record				
INSTRUMENTS USED DURING		INSTRUMENT		M108,VF4-N31
PREPARATION				
Sampl	le ID			
Tested				
Tester				
Record	ded 06/20/24			
EXPIRATION DATE		MM/DD/YY		09/30/24
		MM/DD/YYYY		
Sampl	le ID			
Tested				
Tester				
Record				
Lot # 40	062034			
	5/20/24			

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

1