

Page: 1

Date: 06/18/24 at 9:47 AM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569

Certificate of Analysis / QC Results

Customer PO #:

		Custome	# FO # .	
Packaged Product:	6922			
	CONDUC	IIVITY STANDARD 500 uMHOS		
Test		Target/UOM	Range	Result
PREPARED TO FORMU	LATION ON	YES		YES
FILE		YES		
Sample ID				
Tested	06/18/24			
Tester	RSHE			
Recorded	06/18/24			
ACTUAL CONDUCTIVIT	Y @25°C	500.0	499.9 - 500.1	500.0
		US/CM		
Sample ID				
Tested	06/18/24			
Tester	RSHE			
Recorded	06/18/24			
QCSP30:CONDUCTIVIT	Υ	YES		YES
STD&MEASUREMENT		QCSP30		
Sample ID				
Tested	06/18/24			
Tester	RSHE			
Recorded	06/18/24			
TRACEABLE TO N.I.S.T	. SRM 3191	YES		YES
		YES		
Sample ID				
Tested	06/18/24			
Tester	RSHE			
Recorded	06/18/24			
APPEARANCE AND CO	LOR	CLEAR WATER WHITE		PASS
		LIQUID		
		APPEARANCE		
Sample ID				
Tested	06/18/24			
Tester	RSHE			
Recorded	06/18/24			
INSTRUMENTS USED D	URING	INSTRUMENT		T-229
PREPARATION				
Sample ID				
Tested	06/18/24			
Tester	RSHE			
Recorded	06/18/24			
SHELF LIFE OF PRODU	ICT	SHELF LIFE		1 YEAR
		TIME		
Sample ID				
Tested	06/18/24			
Tester	RSHE			
Recorded	06/18/24	MM/DD 007		00/20/07
EXPIRATION DATE		MM/DD/YY		06/30/25
		MM/DD/YYYY		
Sample ID				
Tested	06/18/24			
Tester	RSHE			
Recorded	06/18/24			

Page: 2

Date: 06/18/24 at 9:47 AM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Lot # 4061898 Made 06/18/24

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager