

Page: 1

Date: 06/18/24 at 4:02 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Packaged Produ	uct: CY157F			
CYANIDE STANDARD 100 PPM				
Test		Target/UOM	Range	Result
PREPARED TO FOR	RMULATION ON	YES		YES
FILE		YES		
Sample	e ID			
Tested	06/18/24			
Tester	SHAR			
Record	ed 06/18/24			
Actual PPM		100.00 PPM	98.00 - 102.00	98.79
Sample	e ID			
Tested	06/18/24			
Tester	SHAR			
Record	ed 06/18/24			
QCSP47:STANDARDIZATION		YES		YES
VARIOUS CYANIDE		QCSP47		
Sample	e ID			
Tested	06/18/24			
Tester	SHAR			
Record	ed 06/18/24			
APPEARANCE AND COLOR		CLEAR LIQUID		PASS
		APPEARANCE		
Sample	e ID			
Tested	06/18/24			
Tester	SHAR			
Record	ed 06/18/24			
INSTRUMENTS USE	ED DURING	INSTRUMENT		P175,BID079,TID41124
PREPARATION				
Sample	e ID			
Tested	06/18/24			
Tester	SHAR			
Record	ed 06/18/24			
SHELF LIFE OF PRODUCT		6 WEEKS TIME		6 WEEKS
Sample	e ID			
Tested	06/18/24			
Tester	SHAR			
Record	ed 06/18/24			
EXPIRATION DATE		MM/DD/YYYY MM/DD/YYYY		07/30/2024
Sample				
Tested	06/18/24			
Tester	SHAR			
Record				
Lot # 40	61730			
	/18/24			



Page: 2

Date: 06/18/24 at 4:02 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager