



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

| Packaged Product: | | T5112-4L TETRAHYDROFURAN HPLC (w/BHT) (4 LITERS) (TX0282-1) | |
|---|---------------------------|--|--------------|
| Test | Target/UOM | Range | Result |
| ASSAY (GC) | MIN 99.9% % | | 99.90% |
| Sample ID | | | |
| Tested | 06/18/24 | | |
| Tester | TBUR | | |
| Recorded | 06/18/24 | | |
| COLOR (APHA) | MAX 10 NUMERICAL | | 10 APHA |
| Sample ID | | | |
| Tested | 06/18/24 | | |
| Tester | TBUR | | |
| Recorded | 06/18/24 | | |
| FILTERED THROUGH A 0.2 UM FILTER | PASS YES | | PASSES TEST |
| Sample ID | | | |
| Tested | 06/18/24 | | |
| Tester | TBUR | | |
| Recorded | 06/18/24 | | |
| FORM | CLEAR LIQUID FORM | | CLEAR LIQUID |
| Sample ID | | | |
| Tested | 06/18/24 | | |
| Tester | TBUR | | |
| Recorded | 06/18/24 | | |
| Identity IR Spectrum | CONFORMS CONFORMS | | CONFORMS |
| Sample ID | | | |
| Tested | 06/18/24 | | |
| Tester | TBUR | | |
| Recorded | 06/18/24 | | |
| PEROXIDE | MAX 0.01% PEROXIDE | | <0.001% |
| Sample ID | | | |
| Tested | 06/18/24 | | |
| Tester | TBUR | | |
| Recorded | 06/18/24 | | |
| REFRACTIVE INDEX | 1.4070 REFRACTIVE INDE | 1.4060 - 1.4080 | 1.4074 |
| Sample ID | | | |
| Tested | 06/18/24 | | |
| Tester | TBUR | | |
| Recorded | 06/18/24 | | |
| STABILIZER (BHT)% | 0.0200 W/W% | 0.0140 - 0.0310 | 0.0220 |
| Sample ID | | | |
| Tested | 06/18/24 | | |
| Tester | TBUR | | |
| Recorded | 06/18/24 | | |



Aqua Solutions, Inc.
 Ph: 281-479-2569
 Fax: 281-479-2790
 P.O. Box 70
 Deer Park, TX 77536
 Website: www.aquasolutions.org

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569

Certificate of Analysis / QC Results

Customer PO # :

| | | | |
|-----------------------|-----------|-------------|---------|
| Titration Acid | 0.20 | 0.00 - 1.00 | 0.86 |
| | UEQ/G | | |
| Sample ID | | | |
| Tested | 06/18/24 | | |
| Tester | TBUR | | |
| Recorded | 06/18/24 | | |
| PERCENT WATER | MAX 0.03% | | 0.0023% |
| | % WATER | | |
| Sample ID | | | |
| Tested | 06/18/24 | | |
| Tester | TBUR | | |
| Recorded | 06/18/24 | | |
| Lot # | 416902 | | |
| Made | 06/17/24 | | |

Quality Manager