Page: 1

Date: 06/18/24 at 3:01 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

_	Iron Standard 0.1 ppm ASTM E394			
Test		Target/UOM	Range	Result
	ED TO FORMULATION O			YES
FILE		YES		
	Sample ID			
	Tested 06/18/24			
	Tester AROS			
	Recorded 06/18/24			
CALCULATED PPM IRON (FE)		0.10	0.09 - 0.11	0.10
		PPM		
	Sample ID			
	Tested 06/18/24			
	Tester AROS			
	Recorded 06/18/24			
APPEARANCE AND COLOR		CLEAR LIQUID		PASS
		APPEARANCE		
	Sample ID			
	Tested 06/18/24			
	Tester AROS			
	Recorded 06/18/24			
NSTRUM	ENTS USED DURING	INSTRUMENT		P94,VF1-Z6
PREPARA	ATION			
	Sample ID			
	Tested 06/18/24			
	Tester AROS			
	Recorded 06/18/24			
EXPIRATION DATE		MM/DD/YYYY		08/23/2024
		MM/DD/YYYY		
	Sample ID			
	Tested 06/18/24			
	Tester AROS			
	Recorded 06/18/24			

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

huland Is