

Page: 1

Date: 06/18/24 at 3:56 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

	CHLORI	NE STANDARD 1000 PPM AS		
Test		Target/UOM	Range	Result
PREPARED TO FORMULATION ON		YES		YES
FILE		YES		
Sample	ID			
Tested	06/18/24			
Tester	SHAR			
Recorde	ed 06/18/24			
PPM CHLORINE (CL2)		1000.0	995.0 - 1005.0	996.8
		PPM		
Sample	ID			
Tested	06/18/24			
Tester	SHAR			
Recorde				
APPEARANCE AND COLOR		CLEAR PALE YELLOW		PASS
		APPEARANCE		
Sample	ID			
Tested	06/18/24			
Tester	SHAR			
Recorde				
Actual Density	24 00/10/24	1.00100	0.99969 - 1.00230	1.00132
Actual Delisity		GM/ML	0.00000 1.00200	1.00102
Sample	ID	OW/WE		
Tested				
	06/18/24			
Tester	SHAR			
Recorde		INICTOLIMENT		MEE DOO DIDOZO TIDAAAAA
INSTRUMENTS USE	D DURING	INSTRUMENT		M55,P69,BID079,TID41124
PREPARATION				
Sample				
Tested	06/18/24			
Tester	SHAR			
Recorde	ed 06/18/24			
EXPIRATION DATE		MM/DD/YY		08/23/24
		MM/DD/YYYY		
Sample	ID			
Tested	06/18/24			
Tester	SHAR			
Recorde	ed 06/18/24			
SHELF LIFE OF PRO	DDUCT	2 MONTHS		2 MONTHS
		TIME		
Sample	ID			
Tested	06/18/24			
Tester	SHAR			
Recorde				
	51756			
_0				
Made 06/	18/24			

Page: 2

Date: 06/18/24 at 3:56 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager