



**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

<b>Packaged Product:</b> ERL114AF			
POTASSIUM HYDROXIDE 0.1 NORMAL IN IPA			
Test	Target/UOM	Range	Result
<b>PREPARED TO FORMULATION ON FILE</b>	YES		YES
Sample ID			
Tested 06/18/24			
Tester SHAR			
Recorded 06/18/24			
<b>SAMPLE COLLECTION &amp; HANDLING PER LAB040</b>	YES		YES
Sample ID			
Tested 06/18/24			
Tester SHAR			
Recorded 06/18/24			
<b>DATE OF SAMPLING FOR TEST PURPOSES</b>	MM/DD/YYYY		06/18/2024
Sample ID			
Tested 06/18/24			
Tester SHAR			
Recorded 06/18/24			
<b>AMBIENT TEMPERATURE DURING SAMPLING DEGC</b>	25.0 DEG C	10.0 - 35.0	21.0
Sample ID			
Tested 06/18/24			
Tester SHAR			
Recorded 06/18/24			
<b>DATE OF TESTING OF SAMPLE</b>	MM/DD/YYYY MM/DD/YYYY		06/18/2024
Sample ID			
Tested 06/18/24			
Tester SHAR			
Recorded 06/18/24			
<b>AMBIENT TEST TEMPERATURE DEG C</b>	20.0 DEG C	15.0 - 30.0	21.0
Sample ID			
Tested 06/18/24			
Tester SHAR			
Recorded 06/18/24			
<b>LOT NUMBER OF ORIGINAL MATERIAL</b>	BULK LOT NUMBER LOT #		4031808
Sample ID			
Tested 06/18/24			
Tester SHAR			
Recorded 06/18/24			
<b>Actual Normality</b>	0.1000 EQ/L	0.0995 - 0.1005	0.0999
Sample ID			
Tested 06/18/24			
Tester SHAR			
Recorded 06/18/24			



**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**

**Certificate of Analysis / QC Results**

**Customer PO # :**

<b>Actual Density</b>	0.79082 GM/ML	0.78970 - 0.79194	0.79169
Sample ID			
Tested	06/18/24		
Tester	SHAR		
Recorded	06/18/24		
<b>N.I.S.T. Traceable to SRM 84 (KHP)</b>	YES YES		YES
Sample ID			
Tested	06/18/24		
Tester	SHAR		
Recorded	06/18/24		
<b>APPEARANCE AND COLOR</b>	CLEAR-SLIGHTLY HAZY LIQUID APPEARANCE		PASS
Sample ID			
Tested	06/18/24		
Tester	SHAR		
Recorded	06/18/24		
<b>INSTRUMENTS USED DURING PREPARATION</b>	LAB EQUIPMENT INSTRUMENT		M55,BID079,PIDM91,TID41124
Sample ID			
Tested	06/18/24		
Tester	SHAR		
Recorded	06/18/24		
<b>EXPIRATION DATE</b>	MM/DD/YYYY MM/DD/YYYY		07/02/2024
Sample ID			
Tested	06/18/24		
Tester	SHAR		
Recorded	06/18/24		
<b>SHELF LIFE OF PRODUCT</b>	1 WEEK TIME		1 WEEK
Sample ID			
Tested	06/18/24		
Tester	SHAR		
Recorded	06/18/24		

**Lot #** 4061708

**Made** 06/18/24

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.



Aqua Solutions, Inc.  
Ph: 281-479-2569  
Fax: 281-479-2790  
P.O. Box 70  
Deer Park, TX 77536  
Website: [www.aquasolutions.org](http://www.aquasolutions.org)

Page: 3

Date: 06/18/24 at 9:48 AM

**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

**Quality Manager**