



**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

<b>Packaged Product:</b> 1528			
BUFFER REFERENCE pH 11.0 (± 0.01)			
Test	Target/UOM	Range	Result
<b>PREPARED TO FORMULATION ON FILE</b>	YES		YES
Sample ID	YES		
Tested 06/20/24			
Tester RSHE			
Recorded 06/20/24			
<b>Actual pH of product</b>	11.000	10.980 - 11.020	11.005
	PH		
Sample ID			
Tested 06/20/24			
Tester RSHE			
Recorded 06/20/24			
<b>QCSP 20 PH CALIBRATION METHODS</b>	YES		YES
	METHOD		
Sample ID			
Tested 06/20/24			
Tester RSHE			
Recorded 06/20/24			
<b>Actual Temp</b>	25.0	24.0 - 26.0	25.0
	DEG C		
Sample ID			
Tested 06/20/24			
Tester RSHE			
Recorded 06/20/24			
<b>APPEARANCE AND COLOR</b>	CLEAR WATER WHITE		PASS
	LIQUID		
	APPEARANCE		
Sample ID			
Tested 06/20/24			
Tester RSHE			
Recorded 06/20/24			
<b>INSTRUMENTS USED DURING PREPARATION</b>	LAB EQUIPMENT		M108,T227
	INSTRUMENT		
Sample ID			
Tested 06/20/24			
Tester RSHE			
Recorded 06/20/24			
<b>PH NIST TRACEABLE TO SRM 187</b>	YES		YES
	YES		
Sample ID			
Tested 06/20/24			
Tester RSHE			
Recorded 06/20/24			
<b>PH NIST TRACEABLE TO SRM 2193</b>	YES		YES
	YES		
Sample ID			
Tested 06/20/24			
Tester RSHE			
Recorded 06/20/24			



**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

<b>TRACEABLE TO NIST SRM 191I AND SRM 191II</b>	YES YES	YES
Sample ID		
Tested	06/20/24	
Tester	RSHE	
Recorded	06/20/24	
<b>EXPIRATION DATE</b>	MM/DD/YY MM/DD/YYYY	06/30/25
Sample ID		
Tested	06/20/24	
Tester	RSHE	
Recorded	06/20/24	
<b>SHELF LIFE OF PRODUCT</b>	SHELF LIFE TIME	1 YEAR
Sample ID		
Tested	06/20/24	
Tester	RSHE	
Recorded	06/20/24	
<b>Lot #</b>	4060553	
<b>Made</b>	06/20/24	

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.



**Quality Manager**