

Page: 1

Date: 06/25/24 at 2:16 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569

Certificate of Analysis / QC Results

Customer PO #:

Packaged Product:	6939			
· ·	CONDUCT	IVITY STANDARD 25.0 uMHOS		
Test		Target/UOM	Range	Result
PREPARED TO FORMULA	TION ON	YES	90	YES
FILE		YES		
Sample ID				
Tested 0	6/25/24			
Tester R	RSHE			
Recorded 0	6/25/24			
CONDUCTIVITY TO 1 DEC	PLACE	25.0	24.0 - 26.0	25.0
		US/CM		
Sample ID				
Tested 0	6/25/24			
Tester R	RSHE			
Recorded 0	6/25/24			
QCSP30:CONDUCTIVITY		YES		YES
STD&MEASUREMENT		QCSP30		
Sample ID				
•	6/25/24			
Tester R	RSHE			
Recorded 0	6/25/24			
TRACEABLE TO N.I.S.T. S	RM 3191	YES		YES
		YES		
Sample ID				
	6/25/24			
Tester R	RSHE			
Recorded 0	6/25/24			
APPEARANCE AND COLO	R	CLEAR WATER WHITE		PASS
		LIQUID		
		APPEARANCE		
Sample ID				
•	6/25/24			
	RSHE			
	6/25/24			
INSTRUMENTS USED DUF	RING	INSTRUMENT		T-234
PREPARATION				
Sample ID				
	6/25/24			
	RSHE			
	6/25/24			
EXPIRATION DATE		MM/DD/YYYY		12/30/2024
		MM/DD/YYYY		
Sample ID				
Tested 0	6/25/24			
Tester R	RSHE			
Recorded 0	6/25/24			
Lot # 4062588	8		·	
Made 06/25/2				
IVIAUC 00/20/2				



Page: 2

Date: 06/25/24 at 2:16 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager