

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO # :

Packaged	Product:	1794			
-		CHLORIN	NE STANDARD 500 PPM AS CL	2	
Test			Target/UOM	Range	Result
	TO FORMUL	ATION ON	YES	*	YES
FILE			YES		
	Sample ID				
	Tested (06/24/24			
	Tester /	AFRE			
		06/24/24			
CHLORINE PPM			500	495 - 505	498
			PPM		
	Sample ID				
	Tested (06/24/24			
	Tester	AFRE			
	Recorded (06/24/24			
APPEARANCE AND COLO		OR	CLEAR PALE YELLOW		PASS
			APPEARANCE		
	Sample ID				
	-	06/24/24			
	Tester /	AFRE			
		06/24/24			
Actual Dens			1.00020	0.99955 - 1.00085	1.00004
	-		GM/ML		
	Sample ID				
	•	06/24/24			
		AFRE			
		06/24/24			
INSTRUMEN	ITS USED DU		INSTRUMENT		M55,BID00488,PIDP54,TID411
PREPARATION					10
	Sample ID				
		06/24/24			
		AFRE			
		06/24/24			
TRACEABLI	E TO NIST SR		YES		YES
			YES		
	Sample ID				
	-	06/24/24			
		AFRE			
		06/24/24			
EXPIRATIO			MM/DD/YY		08/29/24
	, DAIL		MM/DD/YYYY		00,20,24
	Sample ID				
	-	06/24/24			
		AFRE			
		06/24/24			
SHELF LIFE		00/LT/LT	2 MONTHS		2 MONTHS
			TIME		
	Sample ID				
		06/04/04			
		06/24/24 AFRE			
		AFRE 06/24/24			
		00/24/24			



Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO # :

Lot #	4062047
Made	06/24/24

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

thand St

Quality Manager