



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: 6600			
PERCHLORIC ACID TITRANT 0.1 NORMAL ACETOUS			
Test	Target/UOM	Range	Result
Actual Normality	0.1000	0.0995 - 0.1005	0.1001
	EQ/L		
Sample ID			
Tested 06/28/24			
Tester JENC			
Recorded 06/28/24			
QCSP16:STANDARDIZATION OF ACETOUS PERCHL	YES		YES
	QCSP16		
Sample ID			
Tested 06/28/24			
Tester JENC			
Recorded 06/28/24			
N.I.S.T. Traceable to SRM 84	Yes		YES
	YES		
Sample ID			
Tested 06/28/24			
Tester JENC			
Recorded 06/28/24			
PREPARED TO FORMULATION ON FILE	YES		YES
	YES		
Sample ID			
Tested 06/28/24			
Tester JENC			
Recorded 06/28/24			
APPEARANCE AND COLOR	CLEAR WATER WHITE		PASS
	LIQUID		
	APPEARANCE		
Sample ID			
Tested 06/28/24			
Tester JENC			
Recorded 06/28/24			
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		M87,T-246,VF6-W7,BID-02259, PID-M91
Sample ID			
Tested 06/28/24			
Tester JENC			
Recorded 06/28/24			
EXPIRATION DATE	MM/DD/YY		06/30/25
	MM/DD/YYYY		
Sample ID			
Tested 06/28/24			
Tester JENC			
Recorded 06/28/24			
%WATER	0.020-0.050		0.043
	%WATER		
Sample ID			
Tested 06/28/24			
Tester JENC			
Recorded 06/28/24			



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569

Certificate of Analysis / QC Results

Customer PO # :

Actual Density	1.05793	1.05761 - 1.05825	1.05791
	GM/ML		
Sample ID			
Tested	06/28/24		
Tester	JENC		
Recorded	06/28/24		
NORMALITY HAS BEEN REVISED	YES		YES
	YES		
Sample ID			
Tested	06/28/24		
Tester	JENC		
Recorded	06/28/24		
Lot #	4061863		
Made	06/28/24		

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager