

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO # :

Packaged Produc		NITROGEN STANDARD 1.0	DDM AC N	
Test	AMMONIA	Target/UOM	Range	Result
PREPARED TO FORM	ULATION ON	YES		YES
FILE		YES		
Sample II)			
Tested	06/26/24			
Tester	AROS			
Recorded				
CALCULATED PPM NITROGEN		1.0	0.9 - 1.1	1.1
		PPM		
Sample II)			
Tested	06/26/24			
Tester	AROS			
Recorded				
APPEARANCE AND COLOR		CLEAR LIQUID		PASS
		APPEARANCE		
Sample II)			
Tested	06/26/24			
Tester	AROS			
Recorded				
INSTRUMENTS USED DURING		INSTRUMENT		P160,GC,VF2-A14
PREPARATION	Donano			
Sample I	`			
Tested	06/26/24			
Tester	AROS			
Recorded				
EXPIRATION DATE	00/20/24	MM/DD/YYYY		07/31/2024
		MM/DD/YYYY		5115112024
Sample I)			
Tested	06/26/24			
Tester	AROS			
Recorded	06/26/24			
Lot # 4062	567			
	6/24			

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

huland SPA