



**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

<b>Packaged Product:</b> 8820			
SODIUM THIOSULFATE 1.0 NORMAL			
Test	Target/UOM	Range	Result
<b>Actual Normality</b>	1.0000 EQ/L	0.9990 - 1.0010	1.0004
Sample ID			
Tested	06/25/24		
Tester	ENOR		
Recorded	06/25/24		
<b>QCSP10:STANDARDIZATION OF SODIUM THIOSUL</b>	YES QCSP10		YES
Sample ID			
Tested	06/25/24		
Tester	ENOR		
Recorded	06/25/24		
<b>TITRATION BLANK</b>	0.01 MLS	0.00 - 0.02	0.03
Sample ID			
Tested	06/25/24		
Tester	JRIO		
Recorded	06/25/24		
<b>PREPARED TO FORMULATION ON FILE</b>	YES YES		YES
Sample ID			
Tested	06/25/24		
Tester	JRIO		
Recorded	06/25/24		
<b>Actual Density</b>	1.11981 GM/ML	1.11831 - 1.12132	1.11958
Sample ID			
Tested	06/25/24		
Tester	JRIO		
Recorded	06/25/24		
<b>APPEARANCE AND COLOR</b>	CLEAR WATER WHITE LIQUID APPEARANCE		PASS
Sample ID			
Tested	06/25/24		
Tester	JRIO		
Recorded	06/25/24		
<b>INSTRUMENTS USED DURING PREPARATION</b>	INSTRUMENT		M113,M103,T230
Sample ID			
Tested	06/25/24		
Tester	JRIO		
Recorded	06/25/24		
<b>EXPIRATION DATE</b>	MM/DD/YY MM/DD/YYYY		10/30/24
Sample ID			
Tested	06/25/24		
Tester	JRIO		
Recorded	06/25/24		



Aqua Solutions, Inc.  
 Ph: 281-479-2569  
 Fax: 281-479-2790  
 P.O. Box 70  
 Deer Park, TX 77536  
 Website: [www.aquasolutions.org](http://www.aquasolutions.org)

**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

**SHELF LIFE OF PRODUCT**

4 MONTHS  
 TIME

4 MONTHS

Sample ID

Tested 06/25/24

Tester JRIO

Recorded 06/25/24

**Lot #** 4061806

**Made** 06/25/24

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

**Quality Manager**