

Page: 1

Date: 06/25/24 at 1:23 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569

Certificate of Analysis / QC Results

Customer PO #:

Packaged Product	3844			
	Fluorid	e Standard 30.0 ppm Cer		
Test		Target/UOM	Range	Result
PREPARED TO FORMULATION ON		YES		YES
FILE		YES		
Sample ID				
Tested	06/25/24			
Tester	JENC			
Recorded	06/25/24			
CALCULATED PPM FLUORIDE		30.0	29.0 - 31.0	29.6
		PPM		
Sample ID				
Tested	06/25/24			
Tester	JENC			
Recorded	06/25/24			
QCSP45:MEASURING		YES		YES
WATER		QCSP45		
Sample ID				
Tested	06/25/24			
Tester	JENC			
Recorded	06/25/24			
APPEARANCE AND CO		CLEAR LIQUID		PASS
ALL EXITATION AND OC	LON	APPEARANCE		17.00
Sample ID		ALLEADANOL		
Tested	00/05/04			
	06/25/24			
Tester	JENC 06/25/24			
Recorded		INSTRUMENT		VF12-8,M118
INSTRUMENTS USED DURING		INSTRUMENT		VF12-0,IVI110
PREPARATION				
Sample ID				
Tested	06/25/24			
Tester	JENC			
Recorded	06/25/24			
EXPIRATION DATE		MM/DD/YYYY		12/30/2024
		MM/DD/YYYY		
Sample ID				
Tested	06/25/24			
Tester	JENC			
Recorded	06/25/24			
Lot # 4061	916			
Made 06/2				
IVIAUE 9972	·, - ·			



Page: 2

Date: 06/25/24 at 1:23 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager