

Page: 1

Date: 06/25/24 at 1:25 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569

Certificate of Analysis / QC Results

Customer PO #:

Packaged Product	2099381	र		
WETTING AGENT				
Test		Target/UOM	Range	Result
PREPARED TO FORMU	JLATION ON	YES		YES
FILE		YES		
Sample ID				
Tested	06/25/24			
Tester	MMAT			
Recorded	06/25/24			
MANUFATURE LOT#		MANUFACTURE LOT #		7563629
		LOT#		
Sample ID				
Tested	06/25/24			
Tester	MMAT			
Recorded	06/25/24			
APPEARANCE AND CO		CLEAR LIQUID		PASS
		APPEARANCE		
Sample ID				
Tested	06/25/24			
Tester	MMAT			
Recorded	06/25/24			
Actual Density	00,20,2.	0.99832	0.99821 - 0.99842	0.99826
7101000 2011011,		GM/ML		
Sample ID		S, <u>-</u>		
Tested	06/25/24			
Tester	MMAT			
Recorded	06/25/24			
INSTRUMENTS USED		INSTRUMENT		VF4-N32,EF4-OJK
PREPARATION	Domina	into memeri		VI 1102,21 1 3310
Sample ID				
Tested	06/25/24			
Tester	MMAT			
Recorded	06/25/24			
EXPIRATION DATE	00/20/24	MM/DD/YYYY		06/30/2026
EXPINATION DATE		MM/DD/YYYY		00/00/2020
Sample ID				
Tested	06/25/24			
Tester	MMAT			
Recorded	06/25/24			
Lot # 4062.				
Made 06/2				
Mudc	•			



Page: 2

Date: 06/25/24 at 1:25 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager