

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO # :

Packaged	Product:	EQS148F			
· ·		SULFURIC	ACID 0.01 NORMAL		
Test			Target/UOM	Range	Result
	TO FORMULAT	ION ON	YES	9	YES
FILE			YES		
	Sample ID				
	Tested 06/	27/24			
	Tester SH				
		27/24			
Actual Norm	ality		0.0100	0.0095 - 0.0105	0.0099
			EQ/L		
	Sample ID				
	Tested 06/	27/24			
	Tester SH	AR			
	Recorded 06/	27/24			
APPEARANCE AND COLOR			CLEAR LIQUID		PASS
			APPEARANCE		
	Sample ID				
	Tested 06/	27/24			
	Tester SH	AR			
	Recorded 06/	27/24			
INSTRUMEN	ITS USED DURI	NG	INSTRUMENT		P54,P164,M113,EF6-AMD
PREPARATI	ON				
	Sample ID				
	Tested 06/	27/24			
	Tester SH	AR			
	Recorded 06/	27/24			
EXPIRATION DATE			MM/DD/YY		06/30/25
			MM/DD/YYYY		
	Sample ID				
	Tested 06/	27/24			
	Tester SH	AR			
	Recorded 06/	27/24			
SHELF LIFE	OF PRODUCT		SHELF LIFE		1 YEAR
			TIME		
	Sample ID				
	Tested 06/	27/24			
	Tester SH				
		27/24			
Actual Densi	ity		0.99863	0.99843 - 0.99883	0.99860
			GM/ML		
	Sample ID				
	Tested 06/	27/24			
	Tester SH	AR			
	Recorded 06/	27/24			
Lot	# 4062574				
	e 06/27/24	1			
IVIAU		-			



Page: 2 Date: 06/27/24 at 2:49 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results Customer PO # :

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

1 chard Stop

Quality Manager