



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: HON009F			
QC Calibration Std. 1.5 mg/L Cl Solution			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON	YES		YES
FILE	YES		
Sample ID			
Tested 07/18/24			
Tester RIRO			
Recorded 07/18/24			
PPM CHLORIDE	1.50	1.48 - 1.52	1.50
	PPM		
Sample ID			
Tested 07/18/24			
Tester RIRO			
Recorded 07/18/24			
APPEARANCE AND COLOR	CLEAR LIQUID APPEARANCE		PASS
Sample ID			
Tested 07/18/24			
Tester RIRO			
Recorded 07/18/24			
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		P4,VF1-Z13,BOTTLE
Sample ID			
Tested 07/18/24			
Tester RIRO			
Recorded 07/18/24			
EXPIRATION DATE	MM/DD/YYYY		10/30/2024
	MM/DD/YYYY		
Sample ID			
Tested 07/18/24			
Tester RIRO			
Recorded 07/18/24			
Lot #	4071732		
Made	07/18/24		

Quality Manager