

Page: 1

Date: 07/24/24 at 2:14 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569

Certificate of Analysis / QC Results

Customer PO #:

Ammonia Nitrogen Standard 10 PPM as NH3 Test Target/UOM Range Resu	ult
	ult l
PREPARED TO FORMULATION ON YES YES	3
FILE YES	
Sample ID	
Tested 07/24/24	
Tester AROS	
Recorded 07/24/24	
CERTIFIED PPM NH3 10.0 9.0 - 11.0 10.4	4
PPM	
Sample ID	
Tested 07/24/24	
Tester AROS	
Recorded 07/24/24	
APPEARANCE AND COLOR CLEAR LIQUID PAS	S
APPEARANCE	
Sample ID	
Tested 07/24/24	
Tester AROS	
Recorded 07/24/24	
INSTRUMENTS USED DURING INSTRUMENT P61,GC,V	/F6-17
PREPARATION	
Sample ID	
Tested 07/24/24	
Tester AROS	
Recorded 07/24/24	
EXPIRATION DATE MM/DD/YY 10/30/	/24
MM/DD/YYYY	24
Sample ID	
Tested 07/24/24	
Tester AROS	
Recorded 07/24/24 SHELF LIFE OF PRODUCT SHELF LIFE 3 MON	THE
SHELF LIFE OF PRODUCT SHELF LIFE 3 MON ⁻¹ TIME	1115
Sample ID	
Tested 07/24/24	
Tester AROS	
Recorded 07/24/24	
Lot # 4072284	
Made 07/24/24	



Page: 2

Date: 07/24/24 at 2:14 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager