



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: MOT101F			
POTASSIUM HYDROXIDE IN IPA PH ADJUSTED TO 11.3			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION	YES		YES
	YES		
Sample ID			
Tested	07/22/24		
Tester	JRIO		
Recorded	07/22/24		
APPEARANCE AND COLOR	CLEAR-SLIGHT HAZE/IPA		PASS
	ODOR		
	APPEARANCE		
Sample ID			
Tested	07/22/24		
Tester	JRIO		
Recorded	07/22/24		
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		EF4-FE,M106
Sample ID			
Tested	07/22/24		
Tester	JRIO		
Recorded	07/22/24		
EXPIRATION DATE	MM/DD/YY		07/30/25
	MM/DD/YYYY		
Sample ID			
Tested	07/22/24		
Tester	JRIO		
Recorded	07/22/24		
PH TO 3 DECIMAL PLACES PLACES	11.350	11.300 - 11.400	11.371
	PH		
Sample ID			
Tested	07/22/24		
Tester	JRIO		
Recorded	07/22/24		
Lot #	4071804		
Made	07/22/24		

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.