

Page: 1

Date: 07/25/24 at 4:49 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569

Certificate of Analysis / QC Results

Customer PO #:

Package	ed Product:	VUL385F			
		COPPER S	STANDARD 5.0 PPM SOLUT	CION	
Test			Target/UOM	Range	Result
PREPARE	D TO FORMUL	ATION ON	YES		YES
FILE			YES		
	Sample ID				
	Tested	07/25/24			
	Tester	BPHI			
	Recorded	07/25/24			
PPM CU			5.00	4.95 - 5.05	5.00
			PPM		
	Sample ID				
	Tested	07/25/24			
	Tester	BPHI			
		07/25/24			
APPEARANCE AND COLOR			CLEAR LIQUID		PASS
			APPEARANCE		
	Sample ID				
	Tested	07/25/24			
	Tester	BPHI			
		07/25/24			
TRACEABLE TO N.I.S.T. STANDARD		STANDARD	YES		YES
			YES		
	Sample ID				
	Tested	07/25/24			
	Tester	BPHI			
	Recorded	07/25/24			
INSTRUMENTS USED DURING		URING	INSTRUMENT		P-389,GC,VF6-W10
PREPARATION					
	Sample ID				
	Tested	07/25/24			
	Tester	BPHI			
		07/25/24			
Actual Der	nsity		1.00339	1.00220 - 1.00459	1.00350
			GM/ML		
	Sample ID				
		07/25/24			
		BPHI			
		07/25/24			
EXPIRATION DATE			MM/DD/YY		01/30/25
			MM/DD/YYYY		
	Sample ID				
		07/25/24			
		BPHI			
		07/25/24			
Lo	t# 40722	51			
Ma					
IVIG					



Page: 2

Date: 07/25/24 at 4:49 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager