Page: 1

Date: 07/26/24 at 11:38 AM

## Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

	Silica Standard 1 ppm as Si for EZ1035 Silica Analyzer			
Test		Target/UOM	Range	Result
PREPARE	D TO FORMULATION ON	YES	<u> </u>	YES
FILE		YES		
	Sample ID			
	Tested 07/26/24			
	Tester BPHI			
	Recorded 07/26/24			
PPM SILIC	CA (SIO2)	1.000	0.920 - 1.080	0.944
		PPM W/V		
	Sample ID			
	Tested 07/26/24			
	Tester BPHI			
	Recorded 07/26/24			
APPEARANCE AND COLOR		CLEAR WATER WHITE		PASS
		APPEARANCE		
	Sample ID			
	Tested 07/26/24			
	Tester BPHI			
	Recorded 07/26/24			
INSTRUM	ENTS USED DURING	INSTRUMENT		P-83
PREPARA	TION			
	Sample ID			
	Tested 07/26/24			
	Tester BPHI			
	Recorded 07/26/24			
EXPIRATION	ON DATE	MM/DD/YYYY		04/30/2025
		MM/DD/YYYY		
	Sample ID			
	Tested 07/26/24			
	Tester BPHI			
	Recorded 07/26/24			

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Parland St