Page: 1

Date: 07/24/24 at 2:17 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

	1111110	Ammonia as Nitrogen Standard 10.0 ppm Aqueous Solution			
Test		Target/UOM	Range	Result	
PREPARE	D TO FORMULATION O			YES	
FILE		YES			
	Sample ID				
	Tested 07/24/24				
	Tester AROS				
	Recorded 07/24/24				
PPM NITROGEN 2 DECIMAL		10.00	9.00 - 11.00	10.28	
PLACES		PPM			
	Sample ID				
	Tested 07/24/24				
	Tester AROS				
	Recorded 07/24/24				
APPEARANCE AND COLOR		CLEAR WATER WHITE		PASS	
		APPEARANCE			
	Sample ID				
	Tested 07/24/24				
	Tester AROS				
	Recorded 07/24/24				
NSTRUM	ENTS USED DURING	INSTRUMENT		P69,VF2-B12	
PREPARA	ATION				
	Sample ID				
	Tested 07/24/24				
	Tester AROS				
	Recorded 07/24/24				
EXPIRATION DATE		MM/DD/YYYY		10/30/2024	
		MM/DD/YYYY			
	Sample ID				
	Tested 07/24/24				
	Tester AROS				
	Recorded 07/24/24				

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

huland I