



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Test	Target/UOM	Range	Result
Packaged Product: VUL018F			
	HYDROCHLORIC ACID 0.1 NORMAL (MR/THAM)		
PREPARED TO FORMULATION ON FILE	YES		YES
Sample ID	YES		
Tested 07/26/24			
Tester JENC			
Recorded 07/26/24			
SAMPLE COLLECTION & HANDLING PER LAB040	YES		YES
Sample ID	YES		
Tested 07/26/24			
Tester JENC			
Recorded 07/26/24			
DATE OF SAMPLING FOR TEST PURPOSES	MM/DD/YYYY		07/26/2024
Sample ID	MM/DD/YYYY		
Tested 07/26/24			
Tester JENC			
Recorded 07/26/24			
AMBIENT TEMPERATURE DURING SAMPLING DEGC	25.0	10.0 - 35.0	18.3
Sample ID	DEG C		
Tested 07/26/24			
Tester JENC			
Recorded 07/26/24			
DATE OF TESTING OF SAMPLE	MM/DD/YYYY		07/26/2024
Sample ID	MM/DD/YYYY		
Tested 07/26/24			
Tester JENC			
Recorded 07/26/24			
AMBIENT TEST TEMPERATURE DEG C	20.0	15.0 - 30.0	18.3
Sample ID	DEG C		
Tested 07/26/24			
Tester JENC			
Recorded 07/26/24			
NORMALITY TO 4 DEC PLACES	0.1000	0.0995 - 0.1005	0.1001
Sample ID	EQUIV/LITER		
Tested 07/26/24			
Tester JENC			
Recorded 07/26/24			
N.I.S.T. Traceable to SRM 723	YES		YES
Sample ID	YES		
Tested 07/26/24			
Tester JENC			
Recorded 07/26/24			



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

APPEARANCE AND COLOR	CLEAR LIQUID APPEARANCE	PASS
Sample ID		
Tested 07/26/24		
Tester JENC		
Recorded 07/26/24		
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT	GC,T-62,T-28,M55,BID02259,.PIDM91
Sample ID		
Tested 07/26/24		
Tester JENC		
Recorded 07/26/24		
Actual Density	1.00006 GM/ML	0.99999 - 1.00013 1.00008
Sample ID		
Tested 07/26/24		
Tester JENC		
Recorded 07/26/24		
EXPIRATION DATE	MM/DD/YY MM/DD/YYYY	01/30/26
Sample ID		
Tested 07/26/24		
Tester JENC		
Recorded 07/26/24		
SHELF LIFE OF PRODUCT	18 MONTHS TIME	18 MONTHS
Sample ID		
Tested 07/26/24		
Tester JENC		
Recorded 07/26/24		
Lot #	4070858	
Made	07/26/24	

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.



Quality Manager