



**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

<b>Packaged Product:</b> 130-87-019F			
CONDUCTIVITY STANDARD 50,000 Umhos			
Test	Target/UOM	Range	Result
<b>PREPARED TO FORMULATION ON FILE</b>	YES		YES
Sample ID	YES		
Tested 07/12/24			
Tester RSHE			
Recorded 07/12/24			
<b>ACTUAL CONDUCTIVITY @25°C</b>	50000.0	= 50000.0	50000.0
	US/CM		
Sample ID			
Tested 07/12/24			
Tester RSHE			
Recorded 07/12/24			
<b>TRACEABLE TO N.I.S.T. SRM 3191</b>	YES		YES
	YES		
Sample ID			
Tested 07/12/24			
Tester RSHE			
Recorded 07/12/24			
<b>APPEARANCE AND COLOR</b>	CLEAR LIQUID APPEARANCE		PASS
Sample ID			
Tested 07/12/24			
Tester RSHE			
Recorded 07/12/24			
<b>INSTRUMENTS USED DURING PREPARATION</b>	INSTRUMENT		M92, T234
Sample ID			
Tested 07/12/24			
Tester RSHE			
Recorded 07/12/24			
<b>EXPIRATION DATE</b>	MM/DD/YYYY		07/30/2026
	MM/DD/YYYY		
Sample ID			
Tested 07/12/24			
Tester RSHE			
Recorded 07/12/24			
<b>Lot #</b>	4070327		
<b>Made</b>	07/12/24		



Aqua Solutions, Inc.  
Ph: 281-479-2569  
Fax: 281-479-2790  
P.O. Box 70  
Deer Park, TX 77536  
Website: [www.aquasolutions.org](http://www.aquasolutions.org)

Page: 2

Date: 07/12/24 at 10:12 AM

**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

**Quality Manager**