Page: 1

Date: 07/01/24 at 2:44 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Packaged	Product:	3862			
•			E STANDARD 500 PPM (1M	L=0.5MG)	
Test			Target/UOM	Range	Result
PREPARED	TO FORMU	LATION ON	YES		YES
FILE			YES		
	Sample ID				
	Tested	07/01/24			
	Tester	SHAR			
	Recorded	07/01/24			
APPEARANCE AND COLOR		LOR	CLEAR LIQUID		PASS
			APPEARANCE		
	Sample ID				
	Tested	07/01/24			
	Tester	SHAR			
	Recorded	07/01/24			
INSTRUMEN	NTS USED D	URING	INSTRUMENT		M91,T220
PREPARATI	ON				
	Sample ID				
	Tested	07/01/24			
	Tester	SHAR			
	Recorded	07/01/24			
PPM FLUORIDE			500	490 - 510	503
			PPM		
	Sample ID				
	Tested	07/01/24			
	Tester	SHAR			
	Recorded	07/01/24			
EXPIRATIO			MM/DD/YYYY		01/30/2025
	·-		MM/DD/YYYY		2 ./ 3 3 / 2 3 2 3
	Sample ID				
	Tested	07/01/24			
	Tester	SHAR			
	Recorded	07/01/24			
	# 40627				
Mad	e 07/01	./24			

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

huland I