



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: 8754			
SODIUM STANDARD 1.0 PPM (1000PPB) NIST TRACEABLE SOLUTION			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON FILE	YES		YES
Sample ID	YES		
Tested 07/01/24			
Tester MMAT			
Recorded 07/01/24			
Calculated PPM SODIUM	1.00	0.98 - 1.02	1.00
Sample ID	PPM		
Tested 07/01/24			
Tester MMAT			
Recorded 07/01/24			
APPEARANCE AND COLOR	CLEAR LIQUID APPEARANCE		PASS
Sample ID			
Tested 07/01/24			
Tester MMAT			
Recorded 07/01/24			
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		P138VF1-Z42
Sample ID			
Tested 07/01/24			
Tester MMAT			
Recorded 07/01/24			
SHELF LIFE OF PRODUCT	6 MONTHS TIME		6 MONTHS
Sample ID			
Tested 07/01/24			
Tester MMAT			
Recorded 07/01/24			
EXPIRATION DATE	MM/DD/YYYY		01/30/2025
Sample ID	MM/DD/YYYY		
Tested 07/01/24			
Tester MMAT			
Recorded 07/01/24			
Lot #	4070111		
Made	07/01/24		



Aqua Solutions, Inc.
Ph: 281-479-2569
Fax: 281-479-2790
P.O. Box 70
Deer Park, TX 77536
Website: www.aquasolutions.org

Page: 2

Date: 07/01/24 at 11:39 AM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Quality Manager