



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: 9021			
SULFURIC ACID 0.02 NORMAL (BCG/MR/CO3)			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON FILE	YES		YES
Sample ID	YES		
Tested 07/18/24			
Tester AFRE			
Recorded 07/18/24			
SAMPLE COLLECTION & HANDLING PER LAB040	YES		YES
Sample ID	YES		
Tested 07/18/24			
Tester AFRE			
Recorded 07/18/24			
DATE OF SAMPLING FOR TEST PURPOSES	MM/DD/YYYY		07/18/2024
Sample ID	MM/DD/YYYY		
Tested 07/18/24			
Tester AFRE			
Recorded 07/18/24			
AMBIENT TEMPERATURE DURING SAMPLING DEGC	25.0	10.0 - 35.0	20.0
Sample ID	DEG C		
Tested 07/18/24			
Tester AFRE			
Recorded 07/18/24			
DATE OF TESTING OF SAMPLE	MM/DD/YYYY		07/18/2024
Sample ID	MM/DD/YYYY		
Tested 07/18/24			
Tester AFRE			
Recorded 07/18/24			
AMBIENT TEST TEMPERATURE DEG C	20.0	15.0 - 30.0	20.0
Sample ID	DEG C		
Tested 07/18/24			
Tester AFRE			
Recorded 07/18/24			
NORMALITY TO 4 DEC PLACES	0.0200	0.0195 - 0.0205	0.0195
Sample ID	EQUIV/LITER		
Tested 07/18/24			
Tester AFRE			
Recorded 07/18/24			
QCSP61A:STANDARDIZATION ACIDS USING SODI	YES		YES
Sample ID	QCSP61A		
Tested 07/18/24			
Tester AFRE			
Recorded 07/18/24			




Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

APPEARANCE AND COLOR	CLEAR WATER WHITE LIQUID APPEARANCE	PASS
Sample ID		
Tested	07/18/24	
Tester	AFRE	
Recorded	07/18/24	
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT	GC,T-222,M55,BID03828,PIDM 91
Sample ID		
Tested	07/18/24	
Tester	AFRE	
Recorded	07/18/24	
Actual Density	0.99899 GM/ML	0.99884 - 0.99914 0.99898
Sample ID		
Tested	07/18/24	
Tester	AFRE	
Recorded	07/18/24	
EXPIRATION DATE	MM/DD/YYYY MM/DD/YYYY	07/30/2025
Sample ID		
Tested	07/18/24	
Tester	AFRE	
Recorded	07/18/24	
Lot #	4070202	
Made	07/18/24	

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.



Quality Manager