



Aqua Solutions, Inc.
 Ph: 281-479-2569
 Fax: 281-479-2790
 P.O. Box 70
 Deer Park, TX 77536
 Website: www.aquasolutions.org

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: FIS121F Phosphate Standard 25.0 ppm Solution			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON FILE	YES YES		YES
Sample ID			
Tested 07/24/24			
Tester AFRE			
Recorded 07/24/24			
PPM P04 TO 1 DECIMAL PLACE	25.0 PPM	24.5 - 25.5	25.0
Sample ID			
Tested 07/24/24			
Tester AFRE			
Recorded 07/24/24			
APPEARANCE AND COLOR	CLEAR WATER WHITE APPEARANCE		PASS
Sample ID			
Tested 07/24/24			
Tester AFRE			
Recorded 07/24/24			
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		P115,VF1-Z29
Sample ID			
Tested 07/24/24			
Tester AFRE			
Recorded 07/24/24			
EXPIRATION DATE	MM/DD/YYYY MM/DD/YYYY		01/30/2025
Sample ID			
Tested 07/24/24			
Tester AFRE			
Recorded 07/24/24			
Lot #	4072332		
Made	07/24/24		

Richard [Signature]
Quality Manager